

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1598

Name: RITTER2  
Sterilizer ID: V2313228  
10/14/2022 10:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
2:00 173.6 F 3.2  
4:00 180.9 F 7.0  
6:00 203.8 F 11.9  
8:00 217.3 F 6.4  
10:00 230.3 F 11.7  
12:00 232.9 F 8.8  
14:00 239.4 F 10.5  
16:00 245.8 F 13.6  
18:00 258.3 F 20.5  
20:00 269.0 F 27.4  
22:00 271.9 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.5  
1:00 272.7 F 30.0  
2:00 271.6 F 29.5  
3:00 272.8 F 30.1  
4:00 271.9 F 29.5

Min 271.6 F 29.4  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 215.9 F 0.0  
10:00 227.3 F 0.0  
15:00 232.5 F 0.0  
20:00 228.2 F 0.0  
25:00 228.2 F 0.0  
30:00 224.8 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 22:06  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:35  
CYCLE COMPLETE

10/14/2022 11:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_