
Midmark M11 - v1.0.5
Total Cycles: 1103

Name: RITTER2
Sterilizer ID: V2313228
05/12/2022 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.4 F 0.0
2:00 189.6 F 3.5
4:00 195.2 F 8.8
6:00 220.9 F 7.9
8:00 230.2 F 9.8
10:00 235.8 F 8.8
12:00 244.3 F 12.7
14:00 244.6 F 13.0
16:00 257.2 F 19.8
18:00 268.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.7
2:00 271.7 F 29.0
3:00 271.8 F 29.6
4:00 272.5 F 29.7

Min 271.2 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 202.4 F 0.0
10:00 216.0 F 0.0
15:00 222.6 F 0.0
20:00 223.5 F 0.0
25:00 222.2 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:15
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

05/12/2022 11:26 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____