
Midmark M11 - v1.0.5
Total Cycles: 2692

Name: RITTER2
Sterilizer ID: V2313228
07/24/2023 06:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.3 F 0.0
2:00 178.3 F 2.3
4:00 188.7 F 6.4
6:00 206.0 F 10.9
8:00 220.5 F 5.8
10:00 229.8 F 9.9
12:00 234.4 F 9.3
14:00 242.4 F 12.8
16:00 242.2 F 11.9
18:00 254.5 F 18.2
20:00 265.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.4 F 29.5
2:00 271.6 F 29.4
3:00 272.3 F 29.5
4:00 271.5 F 29.3

Min 271.4 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 206.1 F 0.0
10:00 209.2 F 0.0
15:00 207.1 F 0.0
20:00 208.0 F 0.0
25:00 208.7 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:01:25
CYCLE COMPLETE

07/24/2023 07:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____