
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008087
10:12 21/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:55
STERILIZING 2:56
136.0 C 220kPa 2:56
136.1 C 222kPa 3:26
135.9 C 220kPa 3:56
136.2 C 222kPa 4:26
136.2 C 222kPa 4:56
136.4 C 222kPa 5:26
136.4 C 223kPa 5:56
136.4 C 224kPa 6:26
136.4 C 224kPa 6:26
VENTING 6:27
AIR DRYING 7:19
CYCLE COMPLETE 37:20

Digital Signature #
002495C266BE17DB
