
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001400
9:02 09/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 5:16
STERILIZING 6:36
136.0 C 220kPa 6:36
136.3 C 223kPa 7:06
136.5 C 221kPa 7:36
136.4 C 223kPa 8:06
135.8 C 220kPa 8:36
136.2 C 221kPa 9:06
135.7 C 218kPa 9:36
136.5 C 222kPa 10:06
VENTING 10:06
AIR DRYING 11:05
CYCLE COMPLETE 11:08

Digital Signature #
30E4239EF088E984
