
Midmark M11 - v1.0.0
Total Cycles: 2228

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2022 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 184.0 F 3.5
4:00 190.8 F 8.2
6:00 213.0 F 9.8
8:00 230.0 F 8.3
10:00 239.2 F 13.3
12:00 238.4 F 10.2
14:00 236.4 F 8.2
16:00 248.6 F 14.5
18:00 260.4 F 20.9
20:00 270.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.7 F 29.1
2:00 272.3 F 28.5
3:00 271.3 F 28.2
4:00 272.7 F 29.1

Min 271.2 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 235.6 F 0.0
10:00 233.1 F 0.0
15:00 228.8 F 0.0
20:00 231.8 F 0.0
25:00 224.5 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

08/16/2022 11:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____