
Midmark M11 - v1.0.5
Total Cycles: 2867

Name: RITTER2
Sterilizer ID: V2313228
09/12/2023 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 172.6 F 1.9
4:00 174.6 F 5.4
6:00 195.7 F 10.1
8:00 213.3 F 5.4
10:00 226.6 F 8.9
12:00 236.1 F 11.5
14:00 239.7 F 11.4
16:00 238.3 F 10.0
18:00 251.4 F 16.5
20:00 262.7 F 23.1
22:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.8
2:00 271.9 F 29.1
3:00 271.6 F 29.5
4:00 272.7 F 29.7

Min 271.3 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 206.4 F 0.0
10:00 207.3 F 0.0
15:00 209.9 F 0.0
20:00 208.9 F 0.0
25:00 204.2 F 0.0
30:00 200.8 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

09/12/2023 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____