
Midmark M11 - v1.0.5
Total Cycles: 2376

Name: RITTER2
Sterilizer ID: V2313228
05/16/2023 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.4 F 0.0
2:00 188.7 F 3.8
4:00 205.6 F 9.2
6:00 226.9 F 8.4
8:00 235.3 F 11.4
10:00 235.5 F 9.4
12:00 237.3 F 9.4
14:00 249.0 F 15.2
16:00 261.4 F 22.2
18:00 272.5 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.6 F 29.6
2:00 271.2 F 29.3
3:00 272.1 F 29.3
4:00 271.7 F 29.6

Min 271.0 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 227.9 F 0.0
10:00 239.2 F 0.0
15:00 230.9 F 0.0
20:00 225.9 F 0.0
25:00 231.1 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 2:32
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

05/16/2023 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____