
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004464
10:50 03/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 3:58
136.0 C 220kPa 3:58
136.2 C 222kPa 4:28
136.1 C 223kPa 4:58
136.2 C 220kPa 5:28
136.2 C 222kPa 5:58
136.4 C 222kPa 6:28
136.2 C 223kPa 6:58
136.3 C 224kPa 7:28
136.3 C 224kPa 7:28
VENTING 7:28
AIR DRYING 8:23
CYCLE COMPLETE 8:25

Digital Signature #
9675ABA9B6626E72
