
Midmark M11 - v1.0.5
Total Cycles: 2009

Name: RITTER2
Sterilizer ID: V2313228
02/06/2023 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.7 F 0.0
2:00 139.7 F 0.0
4:00 148.0 F 0.0
6:00 187.2 F 1.3
8:00 207.4 F 4.1
10:00 223.3 F 8.1
12:00 233.2 F 13.0
14:00 235.7 F 9.6
16:00 238.3 F 9.4
18:00 243.4 F 11.7
20:00 247.1 F 13.9
22:00 259.9 F 21.0
24:00 271.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.2
2:00 271.5 F 28.5
3:00 272.7 F 29.3
4:00 271.8 F 28.5

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 227.5 F 0.0
10:00 234.0 F 0.0
15:00 228.6 F 0.0
20:00 227.4 F 0.0
25:00 228.8 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:40
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:02:26
CYCLE COMPLETE

02/06/2023 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____