
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006559
11:02 06/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:34
STERILIZING 4:38
136.4 C 221kPa 4:38
136.2 C 222kPa 5:08
136.3 C 223kPa 5:38
136.4 C 225kPa 6:08
136.0 C 221kPa 6:38
136.4 C 222kPa 7:08
136.2 C 221kPa 7:38
136.2 C 223kPa 8:08
136.2 C 223kPa 8:08
VENTING 8:08
AIR DRYING 8:59
CYCLE COMPLETE 17:50

Digital Signature #
024EA009727BD6E4
