
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000887
10:29 27/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:34
STERILIZING 4:50
136.0 C 220kPa 4:50
136.4 C 233kPa 5:20
136.3 C 231kPa 5:50
136.2 C 231kPa 6:20
136.3 C 231kPa 6:50
136.4 C 233kPa 7:20
136.2 C 230kPa 7:50
136.4 C 233kPa 8:20
136.4 C 233kPa 8:20
VENTING 8:20
AIR DRYING 9:18
CYCLE COMPLETE 24:18

Digital Signature #
0906FED8964E77AD
