
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000604
12:11 26/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:52
STERILIZING 5:14
136.0 C 220kPa 5:14
136.4 C 231kPa 5:44
136.3 C 232kPa 6:14
136.2 C 231kPa 6:44
136.3 C 230kPa 7:14
136.2 C 231kPa 7:44
136.4 C 233kPa 8:14
136.3 C 232kPa 8:44
136.3 C 232kPa 8:44
VENTING 8:44
AIR DRYING 9:42
CYCLE COMPLETE 23:05

Digital Signature #
52E0A9337F8C9C56
