

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 006359  
11:02 26/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:57  
PRESSURIZING 1:59  
STERILIZING 2:59  
136.4 C 220kPa 2:59  
136.2 C 223kPa 3:29  
136.2 C 221kPa 3:59  
136.4 C 222kPa 4:29  
136.4 C 224kPa 4:59  
136.5 C 224kPa 5:29  
136.2 C 223kPa 5:59  
136.5 C 224kPa 6:29  
136.5 C 224kPa 6:29  
VENTING 6:29  
AIR DRYING 7:22  
CYCLE COMPLETE 7:34

-----  
Digital Signature #  
3EA9F26FE6DF970C  
-----