
Midmark M11 - v1.0.0
Total Cycles: 2538

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2022 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.2 F 0.0
2:00 136.7 F 1.0
4:00 154.2 F 4.0
6:00 181.7 F 8.0
8:00 205.2 F 12.6
10:00 212.5 F 5.4
12:00 227.3 F 10.1
14:00 233.1 F 8.1
16:00 240.5 F 12.0
18:00 239.1 F 10.1
20:00 252.6 F 16.7
22:00 264.2 F 23.5
24:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.6 F 29.4
2:00 271.9 F 28.7
3:00 271.6 F 28.8
4:00 272.5 F 29.3

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 211.0 F 0.0
10:00 232.1 F 0.0
15:00 229.3 F 0.0
20:00 226.3 F 0.0
25:00 225.2 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:05
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

11/10/2022 11:10 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____