
Midmark M11 - v1.0.0
Total Cycles: 2116

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2022 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 76.7 F 0.0
4:00 116.7 F 0.0
6:00 129.9 F 0.7
8:00 155.4 F 2.9
10:00 179.4 F 6.1
12:00 196.6 F 9.8
14:00 208.5 F 9.5
16:00 221.3 F 5.9
18:00 232.0 F 9.8
20:00 237.2 F 10.0
22:00 241.4 F 11.4
24:00 238.0 F 9.4
26:00 252.1 F 16.3
28:00 263.8 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.6 F 28.3
2:00 271.7 F 28.3
3:00 271.6 F 28.2
4:00 271.7 F 28.1

Min 271.6 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 232.1 F 0.0
Dry Cycle Aborted

FILLING: 0:51
HEATING: 29:56
STERILIZING: 4:00
VENTING: 2:44
DRYING: 2:15
TOTAL CYCLE: 00:39:46

07/12/2022 01:53 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____