
Midmark M11 - v1.0.0
Total Cycles: 3024

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2023 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.2 F 0.0
2:00 178.7 F 3.1
4:00 191.4 F 7.3
6:00 209.9 F 12.6
8:00 222.6 F 7.4
10:00 236.2 F 12.8
12:00 238.1 F 10.3
14:00 235.9 F 8.5
16:00 250.9 F 15.8
18:00 263.4 F 22.9
20:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.8 F 28.3
3:00 271.9 F 28.8
4:00 272.7 F 28.9

Min 271.3 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 216.5 F 0.0
10:00 232.5 F 0.0
15:00 232.3 F 0.0
20:00 227.5 F 0.0
25:00 227.2 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

03/22/2023 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____