

-----  
Midmark M11 - v1.0.0  
Total Cycles: 660

Name: MIDMARK  
Sterilizer ID: V2178198  
02/10/2021 10:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.8 F 0.0  
2:00 164.9 F 2.5  
4:00 179.3 F 6.7  
6:00 237.4 F 12.6  
8:00 220.8 F 8.6  
10:00 236.1 F 8.8  
12:00 242.9 F 13.0  
14:00 243.8 F 11.7  
16:00 257.9 F 19.2  
18:00 270.3 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 272.6 F 28.7  
2:00 271.9 F 28.4  
3:00 272.4 F 28.8  
4:00 272.6 F 28.6  
  
Min 271.7 F 28.4  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.7 F 0.0  
5:00 237.3 F 0.0  
10:00 226.3 F 0.0  
15:00 230.1 F 0.0  
Dry Cycle Aborted

FILLING: 1:02  
HEATING: 19:18  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 19:07  
TOTAL CYCLE: 00:46:24

02/10/2021 11:25 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01