
Midmark M11 - v1.0.0
Total Cycles: 1003

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2021 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.5 F 0.0
2:00 124.9 F 0.3
4:00 133.0 F 1.9
6:00 170.0 F 4.7
8:00 203.9 F 8.4
10:00 210.3 F 12.6
12:00 228.8 F 7.3
14:00 233.5 F 12.6
16:00 237.8 F 9.2
18:00 239.7 F 9.8
20:00 246.1 F 12.9
22:00 257.7 F 19.1
24:00 268.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.7
1:00 272.2 F 28.9
2:00 272.0 F 29.1
3:00 272.4 F 28.8
4:00 271.6 F 28.9

Min 271.6 F 28.6
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 227.1 F 0.0
10:00 220.3 F 0.0
15:00 217.5 F 0.0
20:00 229.9 F 0.0
25:00 234.7 F 0.0
30:00 233.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:54
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:04:00
CYCLE COMPLETE

07/13/2021 01:08 PM
ACCEPTED: X
REJECTED:
Operator: U01