
Midmark M11 - v1.0.5
Total Cycles: 3871

Name: RITTER2
Sterilizer ID: V2313228
08/19/2024 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.6 F 0.0
2:00 78.0 F 0.0
4:00 82.6 F 0.0
6:00 107.5 F 0.7
8:00 138.3 F 3.4
10:00 166.4 F 6.6
12:00 185.8 F 10.0
14:00 207.0 F 9.5
16:00 216.8 F 5.8
18:00 227.4 F 9.4
20:00 239.2 F 13.4
22:00 236.5 F 9.9
24:00 237.7 F 9.6
26:00 246.1 F 13.7
28:00 257.3 F 19.8
30:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.4
2:00 271.6 F 29.3
3:00 272.5 F 29.2
4:00 271.8 F 29.5

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 231.8 F 0.0
10:00 229.7 F 0.0
15:00 214.4 F 0.0
20:00 208.4 F 0.0
25:00 215.9 F 0.0
30:00 194.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

08/19/2024 10:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____