
Midmark M11 - v1.0.5
Total Cycles: 2041

Name: RITTER2
Sterilizer ID: V2313228
02/13/2023 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.8 F 0.0
2:00 205.1 F 3.4
4:00 214.3 F 8.8
6:00 226.9 F 9.0
8:00 233.4 F 10.8
10:00 234.2 F 9.0
12:00 237.8 F 9.7
14:00 249.1 F 15.4
16:00 261.5 F 22.6
18:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.3 F 29.7
2:00 271.5 F 29.4
3:00 272.3 F 29.7
4:00 271.5 F 29.1

Min 271.4 F 29.1
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 235.3 F 0.0
10:00 227.6 F 0.0
15:00 232.4 F 0.0
20:00 230.7 F 0.0
25:00 227.1 F 0.0
30:00 227.4 F 0.0

Drying Complete

FILLING: 1:60
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:36
CYCLE COMPLETE

02/13/2023 01:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____