

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4196

Name: RITTER2  
Sterilizer ID: V2313228  
12/06/2024 06:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.4 F 0.0  
2:00 141.1 F 0.5  
4:00 169.0 F 3.2  
6:00 198.2 F 7.8  
8:00 210.0 F 13.1  
10:00 225.9 F 8.4  
12:00 235.1 F 12.2  
14:00 239.5 F 11.2  
16:00 237.3 F 9.6  
18:00 251.4 F 16.4  
20:00 263.6 F 23.4  
22:00 270.9 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 271.7 F 29.2  
2:00 272.3 F 29.3  
3:00 272.5 F 29.5  
4:00 272.7 F 29.8

Min 271.4 F 29.2  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 195.7 F 0.0  
10:00 194.4 F 0.0  
15:00 193.2 F 0.0  
20:00 191.0 F 0.0  
25:00 190.7 F 0.0  
30:00 189.7 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 22:41  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:01:18  
CYCLE COMPLETE

12/06/2024 07:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_