

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003344  
12:57 30/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:34  
STERILIZING 3:43  
136.4 C 221kPa 3:43  
136.4 C 225kPa 4:13  
136.4 C 224kPa 4:43  
136.5 C 226kPa 5:13  
136.4 C 224kPa 5:43  
136.4 C 225kPa 6:13  
136.2 C 222kPa 6:43  
136.4 C 224kPa 7:13  
136.4 C 224kPa 7:13  
VENTING 7:13  
AIR DRYING 8:03  
CYCLE COMPLETE 21:48

-----  
Digital Signature #  
3471D81E836D3631  
-----