
Midmark M11 - v1.0.0
Total Cycles: 807

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2021 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.5 F 0.0
2:00 167.1 F 2.9
4:00 196.7 F 8.1
6:00 219.3 F 10.3
8:00 231.9 F 10.0
10:00 234.4 F 7.5
12:00 244.5 F 11.9
14:00 244.7 F 12.7
16:00 258.2 F 19.8
18:00 270.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 273.1 F 29.3
2:00 272.7 F 29.5
3:00 271.8 F 29.1
4:00 272.1 F 28.8

Min 270.9 F 28.1
Max 273.3 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 231.5 F 0.0
10:00 227.3 F 0.0
15:00 231.4 F 0.0
20:00 228.1 F 0.0
25:00 230.6 F 0.0
Dry Cycle Aborted

FILLING: 1:42
HEATING: 18:50
STERILIZING: 4:00
VENTING: 2:60
DRYING: 29:49
TOTAL CYCLE: 00:57:21

04/26/2021 11:41 AM
ACCEPTED: X
REJECTED:
Operator: U01