
Midmark M11 - v1.0.5
Total Cycles: 4333

Name: RITTER2
Sterilizer ID: V2313228
02/05/2025 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 73.9 F 0.0
4:00 113.3 F 0.6
6:00 111.2 F 3.0
8:00 147.4 F 6.3
10:00 169.7 F 9.8
12:00 189.0 F 13.0
14:00 207.4 F 4.1
16:00 218.8 F 7.7
18:00 230.1 F 11.8
20:00 231.0 F 7.8
22:00 243.4 F 13.1
24:00 241.9 F 11.6
26:00 254.5 F 18.0
28:00 265.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.3 F 28.9
2:00 272.7 F 29.2
3:00 272.9 F 29.6
4:00 272.3 F 29.4

Min 271.8 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.1 F 0.0
5:00 194.2 F 0.0
10:00 193.1 F 0.0
15:00 196.8 F 0.0
20:00 201.5 F 0.0
25:00 215.0 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:49
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:08:02
CYCLE COMPLETE

02/05/2025 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____