

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1953

Name: RITTER 1  
Sterilizer ID: V2178198  
05/25/2022 02:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.2 F 0.0  
2:00 178.6 F 4.4  
4:00 199.9 F 9.1  
6:00 223.7 F 7.3  
8:00 231.8 F 10.0  
10:00 232.4 F 7.7  
12:00 245.3 F 13.2  
14:00 243.5 F 11.9  
16:00 256.5 F 18.8  
18:00 268.5 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 28.5  
1:00 272.8 F 28.9  
2:00 272.1 F 29.1  
3:00 271.7 F 28.4  
4:00 272.8 F 28.9  
  
Min 271.4 F 28.4  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
5:00 237.7 F 0.0  
10:00 221.6 F 0.0  
15:00 215.6 F 0.0  
20:00 219.7 F 0.0  
25:00 213.2 F 0.0  
30:00 212.3 F 0.0  
  
Drying Complete

FILLING: 1:43  
HEATING: 19:42  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

05/26/2022 06:56 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02