

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3421

Name: RITTER2  
Sterilizer ID: V2313228  
02/29/2024 10:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.8 F 0.0  
2:00 144.4 F 0.7  
4:00 153.0 F 3.2  
6:00 183.6 F 7.4  
8:00 200.6 F 12.4  
10:00 218.3 F 6.3  
12:00 230.9 F 11.0  
14:00 231.7 F 7.9  
16:00 241.8 F 11.6  
18:00 245.2 F 13.2  
20:00 257.4 F 19.8  
22:00 269.1 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.9  
1:00 272.4 F 29.7  
2:00 272.3 F 29.1  
3:00 272.0 F 29.5  
4:00 272.5 F 29.4

Min 271.5 F 28.8  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
5:00 226.3 F 0.0  
10:00 226.0 F 0.0  
15:00 223.0 F 0.0  
20:00 222.7 F 0.0  
25:00 222.1 F 0.0  
30:00 221.5 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 22:57  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:13  
CYCLE COMPLETE

02/29/2024 11:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_