

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000574  
8:23 05/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:24  
STERILIZING 4:49  
136.5 C 222kPa 4:49  
136.2 C 222kPa 5:19  
136.4 C 223kPa 5:49  
136.5 C 223kPa 6:19  
136.4 C 222kPa 6:49  
136.3 C 223kPa 7:19  
136.2 C 217kPa 7:49  
136.4 C 224kPa 8:19  
136.4 C 224kPa 8:19  
VENTING 8:19  
AIR DRYING 9:15  
CYCLE COMPLETE 9:20

-----  
Digital Signature #  
6299FDB932C88560  
-----