

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002084  
12:56 23/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:14  
PRESSURIZING 3:17  
STERILIZING 4:16  
136.4 C 225kPa 4:16  
136.4 C 227kPa 4:46  
136.2 C 224kPa 5:16  
136.4 C 226kPa 5:46  
136.2 C 224kPa 6:16  
136.4 C 225kPa 6:46  
136.4 C 225kPa 7:16  
136.3 C 224kPa 7:46  
136.3 C 224kPa 7:46  
VENTING 7:46  
AIR DRYING 8:40  
CYCLE COMPLETE 38:40  
-----

Digital Signature #  
D5E09EE2C0CFBBDA  
-----