
Midmark M11 - v1.0.5
Total Cycles: 2441

Name: RITTER2
Sterilizer ID: V2313228
05/31/2023 08:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.8 F 0.0
2:00 144.9 F 0.0
4:00 147.3 F 2.0
6:00 169.9 F 5.8
8:00 192.1 F 10.5
10:00 213.6 F 5.4
12:00 227.3 F 10.0
14:00 232.2 F 8.6
16:00 241.3 F 12.7
18:00 241.2 F 11.4
20:00 254.0 F 18.0
22:00 265.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 30.0
2:00 272.4 F 29.6
3:00 271.4 F 29.0
4:00 271.3 F 29.4

Min 271.1 F 28.9
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 204.7 F 0.0
10:00 217.6 F 0.0
15:00 220.3 F 0.0
20:00 225.9 F 0.0
25:00 232.6 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:54
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:00
CYCLE COMPLETE

05/31/2023 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____