
Midmark M11 - v1.0.0
Total Cycles: 1901

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2022 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 184.7 F 3.4
4:00 196.7 F 7.6
6:00 214.3 F 12.5
8:00 225.0 F 6.6
10:00 237.0 F 11.7
12:00 235.1 F 8.4
14:00 241.9 F 10.7
16:00 246.2 F 13.1
18:00 258.4 F 19.7
20:00 269.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.5
2:00 272.6 F 29.0
3:00 271.6 F 28.3
4:00 272.4 F 28.8

Min 271.5 F 28.3
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 233.3 F 0.0
10:00 229.5 F 0.0
15:00 229.8 F 0.0
20:00 230.2 F 0.0
25:00 231.7 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:43
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

05/03/2022 12:52 PM
ACCEPTED: X
REJECTED:
Operator: U02