

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1741

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2022 11:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 184.7 F 3.2  
4:00 196.5 F 7.7  
6:00 216.7 F 12.6  
8:00 229.1 F 8.0  
10:00 241.8 F 13.4  
12:00 240.4 F 11.1  
14:00 238.5 F 9.8  
16:00 253.3 F 17.0  
18:00 265.8 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.4 F 28.8  
2:00 271.8 F 28.9  
3:00 272.1 F 28.6  
4:00 272.1 F 29.0  
  
Min 271.2 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
5:00 244.2 F 0.0  
10:00 212.2 F 0.0  
15:00 228.7 F 0.0  
20:00 215.4 F 0.0  
25:00 228.6 F 0.0  
30:00 210.1 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 19:36  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:21  
CYCLE COMPLETE

03/24/2022 12:36 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02