

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2685

Name: RITTER2  
Sterilizer ID: V2313228  
07/24/2023 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.3 F 0.0  
2:00 164.1 F 0.1  
4:00 175.1 F 3.4  
6:00 194.8 F 6.9  
8:00 208.2 F 10.5  
10:00 215.7 F 6.9  
12:00 226.8 F 8.7  
14:00 235.4 F 13.3  
16:00 237.7 F 11.0  
18:00 236.5 F 9.5  
20:00 249.6 F 15.6  
22:00 260.1 F 21.5  
24:00 269.7 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.4 F 29.8  
2:00 272.1 F 29.3  
3:00 271.2 F 28.9  
4:00 271.7 F 29.6

Min 271.1 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.2 F 0.0  
5:00 228.9 F 0.0  
10:00 212.3 F 0.0  
15:00 195.3 F 0.0  
20:00 194.0 F 0.0  
25:00 198.2 F 0.0  
30:00 208.4 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 25:05  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:04:24  
CYCLE COMPLETE

07/24/2023 10:40 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_