
Midmark M11 - v1.0.0
Total Cycles: 1220

Name: RITTER 1
Sterilizer ID: V2178198
09/30/2021 06:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.2 F 0.0
2:00 89.5 F 0.0
4:00 95.8 F 0.0
6:00 121.2 F 0.0
8:00 152.9 F 0.1
10:00 181.2 F 0.2
12:00 204.8 F 2.3
14:00 221.5 F 5.5
16:00 229.7 F 8.4
18:00 238.2 F 12.0
20:00 232.9 F 7.6
22:00 244.4 F 12.7
24:00 240.6 F 10.4
26:00 236.4 F 8.6
28:00 251.3 F 15.6
30:00 262.9 F 22.2
32:00 272.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.9
1:00 272.4 F 28.6
2:00 271.8 F 27.9
3:00 272.6 F 28.7
4:00 271.8 F 27.9

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 200.2 F 0.0
10:00 196.0 F 0.0
15:00 191.6 F 0.0
20:00 172.1 F 0.0
25:00 174.1 F 0.0
30:00 167.2 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 32:06
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:10:09
CYCLE COMPLETE

09/30/2021 07:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____