
Midmark M11 - v1.0.0
Total Cycles: 2229

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2022 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.8 F 0.0
2:00 173.6 F 3.5
4:00 201.0 F 8.1
6:00 225.4 F 11.7
8:00 227.5 F 8.5
10:00 238.0 F 13.5
12:00 237.3 F 10.5
14:00 237.9 F 8.2
16:00 249.5 F 15.1
18:00 261.1 F 21.8
20:00 271.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.0
2:00 271.2 F 28.6
3:00 272.4 F 29.1
4:00 271.3 F 28.4

Min 271.2 F 28.4
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 210.0 F 0.0
10:00 223.4 F 0.0
15:00 224.4 F 0.0
20:00 227.6 F 0.0
25:00 228.1 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

08/16/2022 12:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____