
Midmark M11 - v1.0.0
Total Cycles: 3259

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2023 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.8 F 0.0
2:00 175.6 F 0.0
4:00 196.0 F 4.2
6:00 211.7 F 9.3
8:00 236.1 F 8.6
10:00 235.3 F 10.2
12:00 236.7 F 8.2
14:00 245.7 F 12.2
16:00 245.5 F 12.8
18:00 257.7 F 19.3
20:00 268.5 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.4 F 28.8
2:00 271.8 F 28.2
3:00 271.4 F 28.3
4:00 272.5 F 28.8

Min 271.1 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 236.3 F 0.0
10:00 230.6 F 0.0
15:00 230.7 F 0.0
20:00 225.7 F 0.0
25:00 228.6 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

05/18/2023 05:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____