
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000775
15:50 28/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:49
STERILIZING 3:53
136.0 C 220kPa 3:53
136.2 C 232kPa 4:23
136.3 C 233kPa 4:53
136.4 C 233kPa 5:23
136.4 C 233kPa 5:53
136.2 C 233kPa 6:23
136.4 C 233kPa 6:53
136.4 C 232kPa 7:23
136.4 C 232kPa 7:23
VENTING 7:23
AIR DRYING 8:19
CYCLE COMPLETE 8:24

Digital Signature #
6162676F73614C78
