

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4330

Name: RITTER 1  
Sterilizer ID: V2178198  
02/23/2024 04:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
2:00 182.6 F 4.1  
4:00 204.5 F 9.5  
6:00 222.1 F 5.7  
8:00 236.8 F 11.0  
10:00 234.7 F 8.5  
12:00 239.1 F 9.7  
14:00 248.7 F 14.5  
16:00 261.9 F 22.0  
18:00 272.2 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.0 F 28.3  
2:00 272.7 F 28.8  
3:00 271.9 F 28.5  
4:00 272.3 F 28.5  
  
Min 271.5 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
5:00 230.8 F 0.0  
10:00 213.3 F 0.0  
15:00 211.4 F 0.0  
20:00 209.1 F 0.0  
25:00 209.3 F 0.0  
30:00 206.4 F 0.0  
Drying Complete

FILLING: 2:12  
HEATING: 18:09  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:28  
CYCLE COMPLETE

02/23/2024 05:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_