
Midmark M11 - v1.0.5
Total Cycles: 1093

Name: RITTER2
Sterilizer ID: V2313228
05/11/2022 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 187.9 F 3.5
4:00 197.6 F 8.5
6:00 215.9 F 9.7
8:00 229.6 F 9.4
10:00 235.4 F 9.3
12:00 244.2 F 13.4
14:00 244.2 F 12.6
16:00 256.8 F 19.3
18:00 267.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.7 F 29.5
2:00 271.9 F 28.8
3:00 271.6 F 29.0
4:00 272.8 F 29.4

Min 271.3 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 211.6 F 0.0
10:00 215.0 F 0.0
15:00 223.3 F 0.0
20:00 232.6 F 0.0
25:00 227.7 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

05/11/2022 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____