

-----  
Midmark M11 - v1.0.5  
Total Cycles: 323

Name: RITTER2  
Sterilizer ID: V2313228  
08/31/2021 06:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.6 F 0.0  
2:00 72.8 F 0.0  
4:00 82.4 F 0.1  
6:00 112.2 F 0.3  
8:00 148.8 F 2.2  
10:00 210.7 F 6.1  
12:00 217.4 F 10.5  
14:00 213.4 F 6.3  
16:00 230.5 F 9.4  
18:00 235.5 F 10.6  
20:00 237.6 F 11.3  
22:00 236.6 F 9.5  
24:00 250.2 F 16.1  
26:00 261.6 F 22.8  
28:00 272.5 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.6  
1:00 272.1 F 29.6  
2:00 272.9 F 30.0  
3:00 272.7 F 30.3  
4:00 271.5 F 29.5

Min 271.5 F 29.4  
Max 273.1 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.6 F 0.0  
5:00 226.1 F 0.0  
10:00 220.3 F 0.0  
15:00 230.4 F 0.0  
20:00 205.4 F 0.0  
25:00 189.9 F 0.0  
30:00 186.4 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 29:02  
STERILIZING: 4:00  
VENTING: 2:50  
DRYING: 30:00  
TOTAL CYCLE: 01:06:54  
CYCLE COMPLETE

08/31/2021 08:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_