
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001489
10:27 10/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:14
PRESSURIZING 7:02
STERILIZING 8:41
136.0 C 220kPa 8:41
136.1 C 219kPa 9:11
135.7 C 217kPa 9:41
136.4 C 222kPa 10:12
136.4 C 223kPa 10:42
136.3 C 222kPa 11:12
136.4 C 222kPa 11:42
136.2 C 220kPa 12:11
VENTING 12:12
AIR DRYING 13:14
CYCLE COMPLETE 14:36

Digital Signature #
94AF23ECF8EFFDAA
