

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3784

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2024 03:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.5 F 0.0  
2:00 152.1 F 2.6  
4:00 167.1 F 6.3  
6:00 189.4 F 10.3  
8:00 208.3 F 6.8  
10:00 222.8 F 8.5  
12:00 234.2 F 13.1  
14:00 234.9 F 10.4  
16:00 235.6 F 8.7  
18:00 247.7 F 15.2  
20:00 259.2 F 21.5  
22:00 269.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.6  
1:00 272.5 F 30.5  
2:00 272.2 F 30.0  
3:00 271.5 F 29.6  
4:00 271.2 F 29.8

Min 271.1 F 29.4  
Max 272.8 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
5:00 195.6 F 0.0  
10:00 189.7 F 0.0  
15:00 186.6 F 0.0  
20:00 183.9 F 0.0  
25:00 181.9 F 0.0  
30:00 179.5 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 23:03  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:02:14  
CYCLE COMPLETE

07/15/2024 04:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_