
Midmark M11 - v1.0.0
Total Cycles: 5250

Name: RITTER 1
Sterilizer ID: V2178198
01/16/2025 02:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.4 F 0.0
2:00 109.5 F 0.0
4:00 126.3 F 0.0
6:00 151.0 F 0.7
8:00 180.4 F 4.0
10:00 196.6 F 7.0
12:00 222.4 F 10.8
14:00 222.0 F 5.4
16:00 231.4 F 9.5
18:00 236.8 F 10.2
20:00 241.5 F 11.7
22:00 238.4 F 9.8
24:00 251.7 F 16.1
26:00 262.6 F 22.4
28:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.0 F 28.8
2:00 272.1 F 28.7
3:00 271.4 F 28.3
4:00 272.4 F 29.1

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 197.8 F 0.0
10:00 195.8 F 0.0
15:00 197.0 F 0.0
20:00 203.5 F 0.0
25:00 238.4 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 28:19
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:06:58
CYCLE COMPLETE

01/16/2025 03:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____