

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1537

Name: RITTER2  
Sterilizer ID: V2313228  
09/26/2022 03:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
2:00 182.0 F 3.8  
4:00 196.5 F 8.7  
6:00 214.3 F 9.7  
8:00 227.3 F 9.1  
10:00 234.7 F 10.1  
12:00 242.0 F 12.8  
14:00 242.2 F 12.0  
16:00 255.1 F 18.8  
18:00 267.0 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.2 F 29.5  
2:00 272.3 F 30.0  
3:00 271.9 F 29.3  
4:00 272.4 F 30.0  
  
Min 271.5 F 29.1  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.7 F 0.0  
5:00 217.2 F 0.0  
10:00 243.1 F 0.0  
15:00 227.9 F 0.0  
20:00 225.4 F 0.0  
25:00 223.0 F 0.0  
30:00 222.4 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:39  
CYCLE COMPLETE

09/26/2022 04:41 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_