
Midmark M11 - v1.0.5
Total Cycles: 435

Name: RITTER2
Sterilizer ID: V2313228
10/08/2021 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 169.4 F 2.4
4:00 172.2 F 6.2
6:00 189.8 F 11.1
8:00 221.2 F 5.7
10:00 225.6 F 10.8
12:00 230.9 F 7.7
14:00 244.3 F 12.7
16:00 244.2 F 12.5
18:00 256.1 F 18.9
20:00 266.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 273.1 F 29.8
2:00 272.3 F 29.6
3:00 272.4 F 29.4
4:00 272.1 F 29.5

Min 270.9 F 28.8
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 236.7 F 0.0
10:00 220.6 F 0.0
15:00 189.0 F 0.0
20:00 177.2 F 0.0
25:00 175.0 F 0.0
30:00 173.2 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

10/08/2021 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____