
Midmark M11 - v1.0.0
Total Cycles: 3620

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2023 05:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.3 F 0.0
2:00 114.8 F 0.0
4:00 140.1 F 0.0
6:00 169.5 F 1.6
8:00 190.9 F 4.8
10:00 209.6 F 8.2
12:00 216.1 F 12.1
14:00 223.9 F 6.2
16:00 234.1 F 10.6
18:00 233.4 F 8.0
20:00 244.1 F 12.9
22:00 242.6 F 11.5
24:00 254.6 F 17.6
26:00 265.2 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 272.4 F 28.9
2:00 271.6 F 28.1
3:00 272.5 F 28.9
4:00 271.9 F 28.3

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 209.5 F 0.0
10:00 209.9 F 0.0
15:00 204.8 F 0.0
20:00 194.5 F 0.0
25:00 193.5 F 0.0
30:00 201.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 27:45
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:06:13
CYCLE COMPLETE

08/04/2023 06:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____