
Midmark M11 - v1.0.5
Total Cycles: 4373

Name: RITTER2
Sterilizer ID: V2313228
02/24/2025 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.7 F 0.0
2:00 125.6 F 1.1
4:00 130.9 F 3.7
6:00 153.0 F 7.2
8:00 180.5 F 11.2
10:00 200.1 F 3.6
12:00 212.3 F 7.9
14:00 225.0 F 12.1
16:00 229.1 F 8.3
18:00 239.3 F 13.3
20:00 240.5 F 11.6
22:00 251.8 F 17.4
24:00 262.1 F 23.4
26:00 271.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.7
1:00 271.8 F 30.3
2:00 272.8 F 30.5
3:00 272.3 F 30.1
4:00 271.7 F 29.8

Min 271.1 F 29.5
Max 272.9 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 214.4 F 0.0
10:00 220.6 F 0.0
15:00 214.4 F 0.0
20:00 200.0 F 0.0
25:00 197.4 F 0.0
30:00 196.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:40
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:05:24
CYCLE COMPLETE

02/24/2025 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____