
Midmark M11 - v1.0.5
Total Cycles: 3859

Name: RITTER2
Sterilizer ID: V2313228
08/14/2024 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.7 F 0.0
2:00 150.6 F 2.3
4:00 165.2 F 5.5
6:00 190.7 F 9.2
8:00 204.9 F 13.4
10:00 219.0 F 7.4
12:00 230.3 F 11.9
14:00 231.3 F 8.7
16:00 239.6 F 11.5
18:00 242.8 F 12.8
20:00 254.8 F 18.8
22:00 265.0 F 25.0
24:00 271.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 271.8 F 29.9
2:00 272.4 F 30.0
3:00 272.4 F 30.4
4:00 272.0 F 29.8

Min 270.9 F 29.2
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 192.5 F 0.0
10:00 190.3 F 0.0
15:00 188.0 F 0.0
20:00 186.0 F 0.0
25:00 184.1 F 0.0
30:00 182.7 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:03:10
CYCLE COMPLETE

08/14/2024 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____