
Midmark M11 - v1.0.0
Total Cycles: 3540

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2023 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.5 F 0.0
2:00 73.0 F 0.0
4:00 82.9 F 0.0
6:00 112.8 F 0.6
8:00 146.9 F 2.0
10:00 170.0 F 4.4
12:00 189.9 F 8.0
14:00 206.3 F 11.7
16:00 214.5 F 4.2
18:00 225.6 F 8.3
20:00 236.1 F 12.6
22:00 235.3 F 9.1
24:00 239.8 F 10.2
26:00 245.9 F 13.2
28:00 257.8 F 19.5
30:00 268.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.2
3:00 272.7 F 29.0
4:00 271.6 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.0 F 0.0
5:00 195.2 F 0.0
10:00 208.0 F 0.0
15:00 210.8 F 0.0
20:00 215.2 F 0.0
25:00 219.4 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:12
CYCLE COMPLETE

07/18/2023 09:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____