
Midmark M11 - v1.0.0
Total Cycles: 1456

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2021 07:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 73.0 F 0.0
4:00 81.9 F 0.0
6:00 121.7 F 0.2
8:00 146.2 F 0.9
10:00 174.8 F 2.5
12:00 195.7 F 4.1
14:00 212.4 F 7.1
16:00 220.9 F 10.4
18:00 227.5 F 8.7
20:00 229.9 F 7.8
22:00 239.3 F 12.1
24:00 236.7 F 9.0
26:00 241.7 F 10.5
28:00 246.3 F 13.0
30:00 257.5 F 19.0
32:00 267.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.5 F 28.8
2:00 272.1 F 28.3
3:00 271.5 F 28.0
4:00 272.4 F 28.8

Min 271.4 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 192.4 F 0.0
10:00 197.2 F 0.0
15:00 196.8 F 0.0
20:00 210.4 F 0.0
25:00 211.5 F 0.0
30:00 216.1 F 0.0

Drying Complete

FILLING: 0:60
HEATING: 33:30
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:11:46
CYCLE COMPLETE

12/15/2021 08:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____